



UK Head Office: Gloucester Road Trading Estate Malmesbury Wiltshire SN16 9JT
+44 (0) 1666 822577 sales@hilditchgroup.com www.hilditchgroup.com

Condition Report

Sale 1829 - 15/01/2019 10:00 - Live Webcast Medical Equipment Auction - PART 2

Sale Date 15-01-2019 10:00

Location: Malmesbury, Wiltshire, UK

Viewing: Friday 11th January 12pm - 5pm, Monday 14th January 10am - 6pm and the morning of the sale from 8:30am

Removal: By 4pm on Tuesday 22nd January 2019 - Late collection charges will be incurred after this date

These reports are compiled by our Medical Engineering Team and are for guidance only - all lots are sold as seen, as per our Conditions of Sale. We recommend that wherever possible you inspect lots prior to purchase.

Lot 1071 Haag Streit Slit Lamp

Condition Report: The slit lamp powers up but the lamp is missing. The binocular assembly is intact with eyepieces in place and the optics appear to be clear (not fully tested). The fixation light, joystick control, chin rest adjustment and table height adjustment are in good working order. The base plate of the storage box is missing and the lid is also loose. Otherwise the unit is in satisfactory physical condition with signs of wear.

Lot 1140 2 x Philips MP70 Intellivue Patient Monitors

Serial No: DE84383003 & DE84382996

Manufactured: 2009

Software: Rev K.21.58

Condition Report: Both critical/cardiac care patient monitor power up and completed the system initialisation. The touch screen, rotary selector knob and soft keys are working fine. The unit do not have parameter module. Function menus are accessible working fine but was not fully tested. Lot are in good physical condition.

Lot 1254 GE Instatrak 3500 Plus Surgical Navigation System with Modem Hard Drive

Serial No: IT35237

Manufactured: 2004

Accessories: LCD/Touch Screen
Keyboard & mouse

Software: IT35237

Condition Report: .

Lot 1264 Blease Frontline 560 Plus Anaesthetic Machine

Condition Report: The anaesthetic machine powers up accordingly and the pressure gauges, flowmeters, O2 flush and basal flow are in in good working order. The unit comes with a built-in O2



UK Head Office: Gloucester Road Trading Estate Malmesbury Wiltshire SN16 9JT
+44 (0) 1666 822577 sales@hilditchgroup.com www.hilditchgroup.com

monitor which appears to be functional but has not been fully tested. Air, O₂, N₂O medical gas hoses are in place and suction regulator with hose and AGSS receiver with hose are also included. Overall the anaesthetic machine is in satisfactory physical condition. The unit does not come with ventilator.

Lot 1265 Blease Frontline 560 Plus Anaesthetic Machine

Condition Report: The anaesthetic machine powers up accordingly and the pressure gauges, flowmeters, O₂ flush and basal flow are in in good working order. The unit comes with a built-in O₂ monitor which appears to be functional but has not been fully tested. Air, O₂, N₂O medical gas hoses are in place and suction regulator with hose and AGSS receiver with hose are also included. Overall the anaesthetic machine is in satisfactory physical condition. The unit does not come with ventilator.

Lot 1266 Blease Frontline 560 Plus Anaesthetic Machine

Condition Report: The anaesthetic machine powers up accordingly and the pressure gauges, flowmeters, O₂ flush and basal flow are in in good working order. The unit comes with a built-in O₂ monitor which appears to be functional but has not been fully tested. Air, O₂, N₂O medical gas hoses are in place and suction regulator with hose and AGSS receiver with hose are also included. Overall the anaesthetic machine is in satisfactory physical condition. The unit does not come with ventilator.

Lot 1268 Haag Streit Slit Lamp

Condition Report: The slit lamp powers up and the lamp and brightness control are in good working order. The binocular assembly is intact with eyepieces in place and the system produces clear images. The joystick control is functional but the roller wheel on one side of the slit lamp is broken. The chin rest adjustment and table height adjustment are in good working order. Overall the unit is in satisfactory physical condition. The mains cable for the power supply is slightly loose causing intermittent power loss.

Lot 1584 Zeiss IOL Master Optical Biometer On Stand | AN 153217 | SN 1026677

Manufactured: 09 2009

Software: 5.4.4.0006

Condition Report: The biometry unit powers up and completes the initialisation process successfully. Display screen, keyboard/touch pad, joystick control and chin rest adjustment are in good working order. The Zeiss application software loads up accordingly and all measurement modes including ALM, KER, ACD and WTW are operational. A basic function check was carried out with the test eye and the measured values found to be corresponding with the test eye at the time of inspection. Test eye is not included. Overall the unit is in reasonable physical condition with minor signs of wear. Keyboard is included. The unit does not come with the original table/trolley.