



UK Head Office: Gloucester Road Trading Estate Malmesbury Wiltshire SN16 9JT  
+44 (0) 1666 822577 sales@hilditchgroup.com www.hilditchgroup.com

## Condition Report

### Sale 1831 - 12/02/2019 10:00:00 - Live Webcast Medical Equipment Auction - PART 2

**Sale Date** 12-02-2019 10:00

**Location:** Malmesbury, Wiltshire, UK

**Viewing:** Friday 8th February 12pm - 5pm, Monday 11th February 10am - 6pm and the morning of the sale from 8:30am

**Removal:** By 4pm on Tuesday 19th February 2019 - Late collection charges will be incurred after this date

**These reports are compiled by our Medical Engineering Team and are for guidance only - all lots are sold as seen, as per our Conditions of Sale. We recommend that wherever possible you inspect lots prior to purchase.**

---

#### Lot 1018 Stonelight Laser Emission with Foot Switch

**Serial No:** TKM06

**Manufactured:** 2008

**Accessories:** Footswitch  
Interlock plug

**Condition Report:** Unit is in good physical condition but the On/Off switch key is not included. Unable to power up the system and test.

---

#### Lot 1025 Alcon Ophthalas 532 Eyelite Laser Photocoagulator

**Serial No:** 0602730301X

**Manufactured:** 10 2006

**Accessories:** Interlock plug  
x2 Safety goggles

**Condition Report:** The unit is in good physical condition. The On/Off switch key is missing therefore unable to fully test the system.

---

#### Lot 1032 Codomics Horizon Multimedia Imager

**Serial No:** 50C00763C

**Condition Report:** System power up and LCD is working but with error "Power Supply Faulty" displayed. Not further test done. Fair physical condition.

---

#### Lot 1038 KLS Martin ME102 Diathermy with Foot Switch

**Serial No:** ME1020100061953

**Accessories:** Monopolar footswitch  
Bipolar footswitch

**Condition Report:** Unit power up with indicators and RF output power selector in good working order. Both bipolar and monopolar output are activated when the foot switches are depressed. RF output power was not measured. Patient plate monitoring circuit are functional. Lot is in



UK Head Office: Gloucester Road Trading Estate Malmesbury Wiltshire SN16 9JT  
+44 (0) 1666 822577 sales@hilditchgroup.com www.hilditchgroup.com

good physical and cosmetic condition.

---

**Lot 1076 Biotek Incu Incubator Analyser**

**Serial No:** 6050153

**Condition Report:** No mains adapter with this tester. Battery is not working as well. Calibration was due: 20/09/2018. Not all accessories are included. Fair physical condition.

---

**Lot 1136 Heidelberg Engineering HRT3 Ophthalmic Laser System with Lenses**

**Serial No:** HRT3-06236

**Manufactured:** 05 2008

**Condition Report:** The computer unit is missing therefore no checks could be carried out. The unit is in reasonable physical condition. Includes 12x lenses (Cyl-1, Cyl-2, Cyl-3, Cyl-4, Cyl-5 and Cyl-6), (Cyl+1, Cyl+2, Cyl+3, Cyl+4, Cyl+5 and Cyl+6).

---

**Lot 1138 Sedia Trolley Computer, HP Printer Monitor and Keyboard**

**Serial No:** 01-95-02-140

**Manufactured:** 2006

**Condition Report:** Hard drive had been removed. LCD and printer power up but not fully tested.

---

**Lot 1139 Ergotron Trolley with Keys**

**Condition Report:** StyleView SV44 Cart with Cisco table microphone, PID/MIDCTS- SX20CODEC camera system, serial no. FTT190402KJ, Precision HD video camera (connection cables missing). UPS and battery are working . LCD not included. Cart is in good physical condition.

---

**Lot 1189 Erbe ICC200 Diathermy Unit with Footswitch | 15076**

**Serial No:** D-1344

**Manufactured:** 2005

**Accessories:** Foot switch

**Condition Report:** Mains On/Off switch is missing. Unable to test unit.

---

**Lot 1203 Fisher & Paykel Neopuff Resusitaire | SN 107194 (Incomplete)**

**Serial No:** 020115000055

**Manufactured:** 01 2002

**Condition Report:** The unit powers up and the indicators, function keys are working fine. The overhead heater and light are working fine when tested in the manual mode. The baby mode was not fully tested, skin temperature probe is not included. The pneumatics system include Neopuff was not fully tested. Oxygen therapy flow meters and panel are missing. Cot tilt is working and mattress is included. Overall the unit is in fair physical condition. Key pad membranes are worn out but functional.

---

**Lot 1204 Fisher & Paykel Neopuff Infant Resusitaire**

**Serial No:** 020215000174

**Manufactured:** 02 2002



UK Head Office: Gloucester Road Trading Estate Malmesbury Wiltshire SN16 9JT  
+44 (0) 1666 822577 sales@hilditchgroup.com www.hilditchgroup.com

**Accessories:** Air & O2 hoses

**Condition Report:** The unit powers up and the indicators, function keys are working fine. The overhead heater and light are working fine when tested in the manual mode. The baby mode was not fully tested, skin temperature probe is not included. The pneumatics system include Neopuff and oxygen therapy flow meters were not fully tested. Cot tilt is working but the mattress is missing. Overall the unit is in fair physical condition. Key pad membranes are worn out but functional.

---

**Lot 1211 Fujifilm FCR XG5000 CR Reader**

**Serial No:** 97050003

**Manufactured:** 10 2009

**Software:** Version V4.7

**Condition Report:** Unit power up with touch screen and soft keys working. Function menus accessible, communication error is displayed due to the system not connected to the network. System was not fully tested. The front cover edge had been chipped off. Otherwise in fair physical condition.

---

**Lot 1214 Fisher & Paykel Neopuff Resusitaire 9001/W130E and 2 x Mattresses**

**Serial No:** 020115000057

**Manufactured:** 01 2002

**Accessories:** O2 hose

**Condition Report:** The unit powers up and the indicators, function keys are working fine. The overhead heater is working fine when tested in the manual mode. The baby mode was not fully tested, skin temperature probe is not included. Over head light is loose and not working. The pneumatics system include Neopuff was not fully tested. Both oxygen therapy flow meters are missing. Cot tilt and mattress are in good condition. Overall the unit is in fair physical condition.

---

**Lot 1215 Fisher & Paykel Neopuff Resusitaire 9001/W130E with 1 x Mattress (Shelf)**

**Serial No:** 020115000051

**Manufactured:** 01 2002

**Accessories:** O2 hose

**Condition Report:** The unit powers up and the indicators, function keys are working fine. The overhead heater and light are working fine when tested in the manual mode. The baby mode was not fully tested, skin temperature probe is not included. The pneumatics system include Neopuff was not fully tested. Both oxygen therapy flow meters are fitted. Cot tilt and mattress are in good condition. Overall the unit is in fair physical condition. Key pad membranes are worn out but functional.

---

**Lot 1224 Uscom Patient Monitor on Roll Stand, Heine EN100 Ophthalmoscope on Roll**

**Serial No:** 100323

**Accessories:** Mains adapter

USCOM 22-12F Doppler transducer, S/N: 70300217

**Software:** Release 1.8.0.0

**Condition Report:**



UK Head Office: Gloucester Road Trading Estate Malmesbury Wiltshire SN16 9JT  
+44 (0) 1666 822577 sales@hilditchgroup.com www.hilditchgroup.com

The USCOM 1A ultrasonic cardiac output monitor power up with the touch screen working fine and function menus accessible. Patient registration and measurements storage in good working order. The Doppler transducer is working fine with good echo signals displayed. System appears to be working fine. The internal battery low warning is displayed during test.

The Heine diagnostic set's ophthalmoscope is working and the power supply is in good working order.

The Propac 242, S/N: GA010353 power up with touch screen, function menus working fine. No accessories included, therefore parameters were not fully tested. Battery is working.

---

**Lot 1266 Maquet Operating Table**

**Serial No:** 00333

**Manufactured:** 1999

**Condition Report:** No hand control available to test the motorised function of the table. The mechanical adjustment of the back rest is not working. Casters brakes brake are working fine. Cushions set is in fair condition. Lot is in fair physical condition.

---

**Lot 1269 Verathon Bladder Scanner with Probe On Stand**

**Serial No:** 00004198

**Manufactured:** 2009

**Accessories:** Scanner probe  
x 1 Battery pack

**Software:** Version U.I.: 2.8.7.6/136

**Condition Report:** The BVI 9400 power up with the battery pack at 50%. Battery charger is not included. The LCD and soft keys are working; function menus are accessible and the probe is oscillating. System was not fully tested. Printer is out of paper and there is a small crack on the probe's plastic housing. Otherwise system is in good physical condition. No. of scan: 1532.

---

**Lot 1278 Erbe ICC 300 Diathermy On Trolley with Foot Switch**

**Serial No:** D-1095

**Manufactured:** 2005

**Accessories:** Bipolar foot switch

**Condition Report:** The unit power up and completed the system self-test. Numerical displays and key pad are in good working order. The RF power outputs and modes selector keys are working fine. Patient monitoring circuit is functional. Unit was tested with an ESU analyser; measured RF power output was slightly off the set values. Unit is working fine. Unit is in good physical condition.

---

**Lot 1291 Erbe ICC 300 Diathermy with Foot Pedals and Trolley**

**Serial No:** F-1065

**Manufactured:** 2004

**Accessories:** Footswitch

**Condition Report:** The unit power up and completed the system self-test. Numerical displays and key pad



UK Head Office: Gloucester Road Trading Estate Malmesbury Wiltshire SN16 9JT  
+44 (0) 1666 822577 sales@hilditchgroup.com www.hilditchgroup.com

are in good working order. The RF power outputs and modes selector keys are working fine. Patient monitoring circuit is functional. Unit was tested with an ESU analyser; measured RF power output correspond to the set values. Unit is working fine. Unit is in good physical condition.

---

**Lot 1311 2 x Dynascope DS7100 Monitors On Stands | 50000533, 50000531**

**Serial No:** 50000533 & 50000531

**Manufactured:** 2008

**Condition Report:** Both monitors power up with the touch screen and function menus working. No accessories included. System parameters were not fully tested. Internal batteries are low. One the units have a few cracks on the casing. Average physical condition.

---

**Lot 1314 BenQ E910 LCD Monitor (Boxed As New) | 2019/01-076, Storz Monitor (Some**

**Manufactured:** 12 2009

**Condition Report:** The Benq 19" LCD monitor power up ok but there is no stand. In good condition.

---

**Lot 1320 GSI Audiostar Pro Audiometer with Display | GS0062853**

**Serial No:** GS0062853

**Manufactured:** 2016

**Accessories:** x 2 Sets of headphone  
Accessories and cables  
User manual  
Keyboard

**Software:** 1.1.5858.16085

**Condition Report:** System power and completed the initialisation. LCD and control console keys are in good working order. Channels Parameter setting, function menus and adjustments are working fine. System was not fully tested. Lot is in good physical and cosmetic condition. Unit was last calibrated 23/03.2016.

---

**Lot 1341 Carestream Direct View CR System with Monitor, Carestream Direct View Remote**

**Serial No:** 11137

**Manufactured:** 10 2008

**Condition Report:** Unit power up and completed the initialisation. The touch screen is working and the function menus are accessible. The computer unit is included and the system appears to be functional but has not been fully tested. The remote operations panel also appears to be working as normal. Nofurther checks carried out.. Fair physical condition.

---

**Lot 1366 5 x Laerdal Suction Units**

**Serial No:** 78171072294, 78181072588, 78171072293, 78520968754 & 78480967338

**Manufactured:** 04 2010

**Condition Report:** All five LSUs power with indicators and vacuum regulator controls in good working order. Suction pump is running fine with good suction power. Unit tested on battery, which is working fine. Good overall condition.  
78520968754 & 78480967338 are with YOM: 2009.



UK Head Office: Gloucester Road Trading Estate Malmesbury Wiltshire SN16 9JT  
+44 (0) 1666 822577 sales@hilditchgroup.com www.hilditchgroup.com

---

**Lot 1383 Topcon RM8000A Auto Refractometer**

**Serial No:** 2396038

**Manufactured:** 2007

**Accessories:** Electric Table

**Software:** Version 1.34

**Condition Report:** The unit power up with CRT monitor and soft keys working fine. The camera images are clear and within focus. Trigger switch is working but system was not fully tested. The table motorised height adjustment is fully functional. Good physical condition.

---

**Lot 1390 Maquet Alphamax Electric Operating Table with Table Attachment and Arm Rest**

Model: 1133.02B2

**Serial No:** 00146

**Manufactured:** 2002

**Accessories:** x 2 head rest (model: 1130.52B0 & 1133.53B0/01)

Arm support side attachment (model No. 1001.62A0, S/N: 02140)

Light weight Leg support (model No. 1130.65B0/98)

Leg support attachment (Model No. 1133.58BC/02.

**Condition Report:** The unit's battery pack is low, needs recharging. The operating table was tested with mains connected; system alarming when using the corded hand control. However, the emergency key pad is working fine with all the motorised positions functional. The cushion set is not complete and the column blue cover missing its screws and not aligned. The following attachment are in in fair condition.

---

**Lot 1415 5 x Laerdal Suction Units**

**Serial No:** 78181072290, 78181072593, 78181072584, 78171072289 & 78480967334

**Manufactured:** 04 2010

**Condition Report:** All five LSUs power with indicators and vacuum regulator controls in good working order. Suction pump is running fine with good suction power. Unit tested on battery, which is working fine. Good overall condition.  
78480967334 is of YOM: 2009.

---

**Lot 1465 Kaps Colposcope SOM 52 | SN 15595**

**Accessories:** Eye piece set WF 12,5x V  
Objective Lens f=300

**Condition Report:** The colposcope light source , bulb and brightness control are working fine. The optical system is clear with good magnification and focus. Filter and fibre optic guide cable is in good working condition. Fair overall condition with one of the casters broken off.

---

**Lot 1470 1 x Draeger Primus Anaesthetic Machine**

**Serial No:** ARSL-0027

**Manufactured:** 2005

**Operating Hours:** 12292

**Accessories:** Air, O2 & N2O hoses  
AGSS hose

**Software:**



UK Head Office: Gloucester Road Trading Estate Malmesbury Wiltshire SN16 9JT  
+44 (0) 1666 822577 sales@hilditchgroup.com www.hilditchgroup.com

Version 02.03.00

**Condition Report:** The anaesthetic machine pipeline pressure gauges, flowmeters, O2 flush and anti-hypoxic safety feature are working fine. Low pressure O2 alarm is working. The ventilator power with LCD, rotary knob, key pad and function menus are working fine. All ventilator modes (Volume mode, Pressure Mode, Pressure Support & Man Spont) are in good working order. CO2 Absorber canister is not included. Breathing assembly and piston are in good working order. Flow sensors & O2 sensor are working fine. System was tested with medical gas and test lung, good ventilation cycles were delivered. The set tidal volume and breath frequency correspond to the resulting minute volume. The backup battery is working. The Primus anaesthetic machine is in good overall condition. O2 & N2O pin-index yoke is fitted at the back. Examination lamp is working.

**Lot 1471 Maquet Alphastar Operating Table (No Head Cushion)**

Model No. 1133.02B0

**Serial No:** 00317

**Manufactured:** 2002

**Condition Report:** The Alphamaxx table power up and all the motorised positions are working when operated with the emergency key pad. Casters and brake are working fine. There is no hand control and the battery needs charging. . System work fine when connected to mains. The cushion set is nor complete. Unit is in fair overall condition.

---

**Lot 1472 Draeger Fabius Anaesthetic Machine Including Absorber**

**Serial No:** ARSK-0107

**Manufactured:** 2002

**Accessories:** Air, O2 & N2o hoses  
AGSS hose  
Cenral vacuum regulator and hose

**Software:** Version 02.80

**Condition Report:** The anaesthetic machine pipeline pressure gauges, flowmeters, O2 flush and anti-hypoxic safety feature are working fine. Low pressure O2 alarm is working. The ventilator power with LCD, rotary knob, key pad and function menus are working fine. Both ventilator modes (IPPV & Man Spont) are in good working order. Breathing assembly with CO2 Absorber canister and piston are in good condition. Flow sensors & O2 sensor are working fine. System was tested with medical gas and test lung, good ventilation cycles were delivered. The set tidal volume and breath frequency correspond to the resulting minute volume. The backup battery is working. The Fabius anaesthetic machine is in good overall condition. O2, N2O & Air pin-index yoke is fitted at the back.

---

**Lot 1473 Draeger Fabius Anaesthetic Machine**

**Serial No:** ARSK-0033

**Manufactured:** 2002



UK Head Office: Gloucester Road Trading Estate Malmesbury Wiltshire SN16 9JT  
+44 (0) 1666 822577 sales@hilditchgroup.com www.hilditchgroup.com

**Accessories:** Air, O2 & N2O hoses  
AGSS Hose  
Central vacuum Hose & Regulator

**Condition Report:** The anaesthetic machine pipeline pressure gauges, flowmeters, O2 flush and anti-hypoxic safety feature are working fine. Low pressure O2 alarm is working. The Fabius anaesthetic machine is in good overall condition. O2, N2O & Air pin-index yoke is fitted at the back.

---

**Lot 1474 Draeger Fabius Anaesthetic Machine**

**Serial No:** ARSK-0032

**Manufactured:** 2002

**Accessories:** Air, O2 & N2o Hoses  
AGSS hose  
Central vacuum regulator and hose

**Condition Report:** The anaesthetic machine pipeline pressure gauges, flowmeters, O2 flush and anti-hypoxic safety feature are working fine. Small internal leak can be heard when N2O gas hose is connected. Low pressure O2 alarm is not working. The Fabius anaesthetic machine is in good overall condition. O2, N2O & Air pin-index yoke is fitted at the back.

---

**Lot 1475 Acime Ambu Ambulatory Care Chair**

**Serial No:** FA-16946

**Manufactured:** 03 2017

**Accessories:** Corded Hand control  
Mattress

**Condition Report:** All the motorised positions are working. The backup battery and hand control are fully functional. The side rails and chair frame are in good physical condition. There are a couple of marks on the mattress.

---

**Lot 1480 Nellcor Puritan 840 Ventilator**

**Serial No:** 3510080209 (BDU)  
0817090241 (GUI)

**Manufactured:** 2008

**Accessories:** Air & O2 hoses

**Condition Report:** The 840 ventilator system power up and completed the initial power on tests. GUI's LCD/touch screen key pad and rotary selector knob working fine. The function menus are accessible and the ventilator modes are selectable. Exp. filter is not included. System was not fully tested with medical gases. Unit is in good physical condition. Back up battery is working fine.

---

**Lot 1481 Draeger Photo-Therapy 4000 Lamp**

**Serial No:** ARMK 0023

**Condition Report:** Unit power up with therapy lamps working fine. Lot is in fair overall condition.

---





UK Head Office: Gloucester Road Trading Estate Malmesbury Wiltshire SN16 9JT  
+44 (0) 1666 822577 sales@hilditchgroup.com www.hilditchgroup.com

---

**Lot 1482 Zeiss Microscope OPMI 99**

**Serial No:** 245936

**Condition Report:** Light source, bulb, intensity selector and fibre optic guide cable are working. Fair optic system with good magnification. One of the eye pieces cap is missing. One The lock screw for the support arm is missing as well.

---

**Lot 1498 4 x LSU Suction Pumps**

**Serial No:** 78240443051, 78110439749, 78440335448 & 784602247242

**Manufactured:** 2004

**Condition Report:** All five LSUs power with indicators and vacuum regulator controls in good working order. Suction pump is running fine with good suction power. Unit tested on battery, which is working fine. Good overall condition.  
LSU with serial No. 78440335448 is YOM: 2003 & 784602247242 YOM: 2002.

---

**Lot 1530 ERBE ICC200 Diathermy with Foot Pedals**

**Serial No:** D-1004

**Accessories:** Footswitch

**Condition Report:** The unit power up and completed the system self-test. Numerical displays and key pad are in good working order. The RF power outputs and modes selector keys are working fine. Patient monitoring circuit is functional. Unit was tested with an ESU analyser; measured RF power output correspond to the set values. Unit is working fine. Unit is in fair physical condition with some of the paint work peeled off on the top cover.

---

**Lot 1562 5 x Laerdal Suction Units**

**Serial No:** 78181072577, 78171072300, 78171072284, 78410223490 & 78510968440

**Manufactured:** 01 2010

**Condition Report:** All five LSUs power with indicators and vacuum regulator controls in good working order. Suction pump is running fine with good suction power. Unit tested on battery, which is working fine. Good overall condition.  
LSU with serial No. 78410223490 is YOM: 2002 & 78510968440 YOM: 2009, with a crack at the back cover.

---

**Lot 1568 5 x Laerdal Suction Units**

**Serial No:** 78171072285, 78171072267, 78181072568, 78321188423 & 78101294780

**Manufactured:** 04 2010

**Condition Report:** All five LSUs power with indicators and vacuum regulator controls in good working order. Suction pump is running fine with good suction power. Unit tested on battery, which is working fine. Good overall condition.  
LSU with serial No. 78321188423 & is YOM: 2011 & 78101294780 YOM: 2012.

---

**Lot 1606 Erbe ICC 350 Diathermy with Foot Switches On Trolley**

**Serial No:** F-1004

**Accessories:** Bipolar foot switch

**Condition Report:** The unit power up and completed the system self-test. Numerical displays and key pad



UK Head Office: Gloucester Road Trading Estate Malmesbury Wiltshire SN16 9JT  
+44 (0) 1666 822577 sales@hilditchgroup.com www.hilditchgroup.com

are in good working order. The RF power outputs and modes selector keys are working fine. Patient monitoring circuit is functional. Unit was tested with an ESU analyser; measured RF power output correspond to the set values, for both monopolar and bipolar modes. Unit is working fine and in fair physical condition.

---

**Lot 1607 5 x Laerdal Suction Units**

**Serial No:** 78181072563, 78181072592, 78181072578, 78410113492 & 78210783390

**Manufactured:** 05 2010

**Condition Report:** All five LSUs power with indicators and vacuum regulator controls in good working order. Suction pump is running fine with good suction power. Unit tested on battery, which is working fine. Good overall condition.  
LSU with serial No. 78410113492 YOM: 2002 & 78210783390 YOM: 2007.

---

**Lot 1620 5 x Laerdal Suction Units**

**Serial No:** 78171072292, 78171072282, 78171072296 & 78410223495

**Manufactured:** 04 2010

**Condition Report:** All five LSUs power with indicators and vacuum regulator controls in good working order. Suction pump is running fine with good suction power. Unit tested on battery, which is working fine. Good overall condition.  
LSU with serial No. 78410223495 YOM: 2002.

---